

# User seal checks

Half-face reusable respirators | cartridge

## NEGATIVE AND POSITIVE PRESSURE




**Cover cartridge inlets with palms**

Breathe **in** gently

Respirator should slightly **collapse**

**If air leaks in** around the face:



**Adjust straps or  
Reposition face piece**




**Cover exhalation valve with palm**

Breathe **out** gently

Respirator should slightly **expand**

**If air leaks out** around the face:



**REPEAT TEST UNTIL SEAL IS ACHIEVED**